

AURELIA® Zero

Examination Gloves



| | | | |
|-------|------|-------|----------|
| 92775 | (XS) | ø 200 | 10 x 200 |
| 92776 | (S) | ø 200 | 10 x 200 |
| 92777 | (M) | ø 200 | 10 x 200 |
| 92778 | (L) | ø 200 | 10 x 200 |
| 92779 | (XL) | ø 200 | 10 x 200 |



Nitrile



Powder Free



Ice Blue



3.5+/- 0.2 g



Finger-Textured




5 Years

Quality Ice Blue Accelerator-Free Nitrile examination gloves that offer increased tensile strength and protection over Latex and Vinyl.

Thin Nitrile examination gloves specifically designed to offer increased comfort, flexibility & sensitivity.

- 3.2mil Powder-Free Accelerator-Free Nitrile - Non-Sterile
- Beaded cuff
- Ambidextrous
- Latex Free
- AQL 1.5

Sizes & Dimensions



| | | |
|-----------|-------------|-------------|
| XS | 70 ± 10 mm | min. 240 mm |
| S | 80 ± 10 mm | min. 240 mm |
| M | 95 ± 10 mm | min. 240 mm |
| L | 110 ± 10 mm | min. 240 mm |
| XL | ≥ 110 mm | min. 240 mm |

Single wall thickness



- ① min. 0.10 mm
- ② min. 0.06 mm
- ③ min. 0.05 mm

Physical properties

| | Unaged | Aged |
|-------------------|---------------|---------------|
| Tensile Strength | min. 14.0 MPa | min. 14.0 MPa |
| Elongation | min. 500 % | min. 400 % |
| Strength at Break | min. 6N | min. 6N |

Quality Signs & Certificates

- 89/686/EEC, 93/42/EWG
- ISO 9001:2008, ISO 13485, CMDCAS 13485 (SGS)
- EN 455 1-4
- EN 420 : 2003 + A 1 : 2009
- EN 374 1,2 & 3, EN 388
- EN 1186: Food Contact Approved (SGS)

