

MICRO-TOUCH® NITRILE ACCELERATOR-FREE

Nitrile Powder-Free
Non-Sterile Examination Glove



PRODUCT DESCRIPTION

Material	Nitrile
Colour	Indigo
Shape	Ambidextrous
Cuff	Beaded
External surface	Textured
Internal surface	Chlorinated

PHYSICAL PROPERTIES

Thickness (single) Typical average values (mm)	Finger	0.120
	Palm	0.080
	Cuff	0.060
Strength (average values)	Before ageing	After ageing
Elongation at break (%)	550	588
Force at break (N)	7.8	8.0
Minimum length (mm)	240	
Shelf life	3 years	
Storage instructions	Keep out of direct sunlight. Store in a cool and dry place. Keep away from sources of ozone or ignition.	
Packaging	10 x 100 pieces / 1000 pieces per case	

PRODUCT REFERENCES

Size / product code	X-Small 5-6	Small 6-7
	700101	700102
	Medium 7-8	Large 8-9
	700103	700104
	X-Large 9-10	
	700105	

MANUFACTURING AND SAFETY STANDARDS

AQL (Pin-Holes)	Waterleak tested Conform European Norm EN 455-1: Inspection Level I AQL 1.5
Protein content	Contains no latex proteins
Primary skin irritation	Not considered a primary irritant as per FHSA Regulation guidelines 16 CFR 1500
Skin sensitization	No evidence of delayed dermal contact sensitization as per ASTM F-720
Viral penetration	Passes ASTM F-1671 using PhiX174
Sterilization	Not applicable
CE-classification Compliances	Class I CE mark EN 455 parts 1, 2, 3 and 4 EN 374 parts 1, 2 and 3 (conform with the performance requirements) ISO 9001

KEY FEATURES AND ADVANTAGES

- Accelerator-Free
- Allergy prevention
- Safety and chemical resistance
- Latex-like comfort and ease-of-use

Ansell Healthcare Europe N.V.

Riverside Business Park Block J Boulevard International 55 B-1070 Brussels, Belgium
Tel.: +32 (0) 2 528 74 00 Fax: +32 (0) 2 528 74 01 Fax Customer Service: +32 (0) 2 528 74 03
<http://www.ansell.eu> E-Mail: info@ansell.eu

Ansell (UK) Limited

30-32 Mariner Lichfield Road Industrial Estate Tamworth Staffordshire B79 7UL, United Kingdom
Phone: +44 (0) 1564 711 802 Fax: +44 (0) 1564 711 344
<http://www.ansell.eu> email: infouk@eu.ansell.com
Issue date: 14/09/2012

Revision N°: 3

Print date: 14/09/2012



ISO 9001 Certificate
Number FM 40130